

DECLARATION OF CONFORMITY

Manufacturing Company

Monark Exercise AB
Box 6
SE-780 50 VANSBRO
Sweden

Monark Exercise AB
Kroons Väg 1
SE-780 50 VANSBRO
Sweden

MODEL NO.	ART. NO.	CLASS OF THE DEVICES
CardioCare 827 E	90827	Class IIa
Ergomedic 828 E	90828	Class IIa
Ergomedic 831E	90831	Class IIa
Ergomedic 839 E	90839	Class IIa
Ergomedic 874 E	90874	Class IIa
Ergomedic 891E	90891	Class IIa
Ergomedic 891E Wingate Option	90891-2	Class IIa
Ergomedic 894 E	90894	Class IIa

INTENDED USE

For tests, training and research to improve performance and health in sports

We the undersigned hereby declare under our sole responsibility that the medical devices specified above conform to the applicable essential requirements listed in Annex I and meet the provisions stated in Annex V of the EC Directive MDD93/42/EEC, the Swedish law 1993:584 and the Swedish Directive LVFS 2003:11. The quality system has been examined by SP Technical Research Institute of Sweden, notified body No: 0402, as attested in EC Certificate No: 15 36 01. We hereby also confirm that all items comply with the criteria in: IEC 60601-1 and IEC 60601-1-2. Procedure of conformity assessment is according to Annex V.

Vansbro 2013-04-17



Andreas Hedbom
Manager Director