

DECLARATION OF CONFORMITY

SPORTS & MEDICAL

MANUFACTURING COMPANY

Monark Exercise AB
Box 6 / Kroons väg 1
SE-780 50 VANSBRO
Sweden

MODEL NAME	ART. NO.	CLASS OF THE DEVICES
CardioCare 827 E	90827	Class IIa
Ergomedic 828 E	90828	Class IIa
Ergomedic 874 E	90874	Class IIa
Ergomedic 891E	90891	Class IIa
Ergomedic 891E Wingate Option	90891-2	Class IIa
Ergomedic 894 E	90894	Class IIa
Monark LC4	LC4-12	Class IIa
Monark LT2	LT2-12	Class IIa
LC6	LC6-13 novo	Class IIa
LC7 TT novo	LC7 TT	Class IIa
Compact Rehab 871 E	90871	Class IIa
Rehab Trainer 881E	90881	Class IIa
Cardio Care 927 E	90927	Class IIa
Ergomedic 939 E	90939 novo	Class IIa
Monark 928 E	90928-12	Class IIa
RT2	RT2-13	Class IIa
RC4	RC4-12	Class IIa
RC6	RC6 novo	Class IIa

INTENDED USE

For tests, training, rehabilitation and research to improve people's health

We, the undersigned, hereby declare under our sole responsibility that the medical devices specified above conform to the applicable essential requirements listed in Annex I and meet the provisions stated in Annex V of the EC Directive MDD93/42/EEC, the Swedish law 1993:584 and the Swedish Directive LVFS 2003:11. The quality system has been examined by SP Technical Research Institute of Sweden, notified body No: 0402, as attested in EC Certificate No: 15 36 01. We hereby also confirm that all items comply with the criteria in: IEC 60601-1 and IEC 60601-1-2. Procedure of conformity assessment is according to Annex V.

Vansbro 2018-06-12



Andreas Hedbom
Manager Director